


**Quantitative determination of alanine aminotransferase
GPT (ALT)
IVD**

Store at 2-8°C

PRINCIPLE OF THE METHOD

Alanine aminotransferase (ALT) or Glutamate pyruvate transaminase (GPT) catalyses the reversible transfer of an amino group from alanine to α -ketoglutarate forming glutamate and pyruvate. The pyruvate produced is reduced to lactate by lactate dehydrogenase (LDH) and NADH:



The rate of decrease in concentration of NADH, measured photometrically, is proportional to the catalytic concentration of ALT present in the sample¹.

CLINICAL SIGNIFICANCE

The ALT is a cellular enzyme, found in highest concentration in liver and kidney.

High levels are observed in hepatic disease like hepatitis, diseases of muscles and traumatism, its better application is in the diagnosis of the diseases of the liver.

When they are used in conjunction with AST aid in the diagnosis of infarcts in the myocardium, since the value of the ALT stays within the normal limits in the presence of elevated levels of AST^{1,4,5}.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

REAGENTS

R 1 Buffer	TRIS pH 7.8	100 mmol/L
	Lactate dehydrogenase (LDH)	1200 U/L
	L-Alanine	500 mmol/L
R 2 Substrate	NADH	0,18 mmol/L
	α -Ketoglutarate	15 mmol/L

PRECAUTIONS

R1: H290-May be corrosive to metals.

Follow the precautionary statements given in MSDS and label of the product.

PREPARATION

Working reagent (WR):

Mix: 4 vol. (R1) Buffer + 1 vol. (R2) Substrate

Stability: 21 days at 2-8°C or 72 hours at room temperature (15-25°C).

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.

Do not use reagents over the expiration date.

Signs of reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 340 nm < 1,00.

ADDITIONAL EQUIPMENT

- Spectrophotometer or colorimeter measuring at 340 nm.
- Thermostatic bath at 25°C, 30°C ó 37°C (\pm 0,1°C).
- Matched cuvettes 1,0 cm light path.
- General laboratory equipment.

SAMPLES

Serum or plasma¹: Stability 7 days at 2-8°C.

PROCEDURE

- Assay conditions:
Wavelength: 340 nm
Cuvette: 1 cm light path
Constant temperature: 25°C / 30°C / 37°C
- Adjust the instrument to zero with distilled water or air.
- Pipette into a cuvette:

WR (mL)	1,0
Sample (μ L)	100
- Mix, incubate for 1 minute.
- Read initial absorbance (A) of the sample, start the stopwatch and read absorbances at 1 minute intervals thereafter for 3 minutes.
- Calculate the difference between absorbances and the average absorbance differences per minute ($\Delta A/\text{min}$).

CALCULATIONS

$$\Delta A/\text{min} \times 1750 = \text{U/L of ALT}$$

Units: One international unit (IU) is the amount of enzyme that transforms 1 μ mol of substrate per minute, in standard conditions. The concentration is expressed in units per litre of sample (U/L).

Temperature conversion factors

To correct results to other temperatures multiply by:

Assay temperature	Conversion factor to		
	25°C	30°C	37°C
25°C	1,00	1,32	1,82
30°C	0,76	1,00	1,39
37°C	0,55	0,72	1,00

QUALITY CONTROL

Control sera are recommended to monitor the performance of assay procedures: CONTROL Normal and Pathologic.

If control values are found outside the defined range, check the instrument, reagents and technique for problems.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES^{4,5}

	25°C	30°C	37°C
Men	up to 22 U/L	29 U/L	40 U/L
Women	up to 18 U/L	22 U/L	32 U/L

Normal newborns have been reported to show a reference range of up to double the adult, attributed to the neonate's hepatocytes. These values decline to adult levels by approximately 3 months of age.

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Measuring range: From *detection limit* of 0 U/L to *linearity limit* of 400 U/L.

If the results obtained were greater than linearity limit, dilute the sample 1/10 with NaCl 9 g/L and multiply the result by 10.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
Mean (U/L)	42,0	116	41,1	115
SD	0,47	0,42	0,76	1,61
CV (%)	1,11	0,36	1,85	1,40

Sensitivity: 1 U/L = 0,00052 ΔA / min.

Accuracy: Results obtained using Audit Diagnostics reagents (y) did not show systematic differences when compared with other commercial reagents (x).

The results obtained using 50 samples were the following:

Correlation coefficient (r^2): 0,99597.

Regression equation: $y=1,1209x + 1,390$.

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

Anticoagulants currently in use like heparin, EDTA, oxalate and fluoride do not affect the results. Hemolysis interferes with the assay¹.

A list of drugs and other interfering substances with ALT determination has been reported^{2,3}.

NOTES

Audit Diagnostics has instruction sheets for several automatic analyzers.

BIBLIOGRAPHY

- Murray R. Alanine aminotransferase. Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 1088-1090.
- Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
- Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
- Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
- Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.