




Quantitative determination of triglycerides IVD

Store at 2-8°C

PRINCIPLE OF THE METHOD

Sample triglycerides incubated with lipoproteinlipase (LPL), liberate glycerol and free fatty acids. Glycerol is converted to glycerol-3-phosphate (G3P) and adenosine-5-diphosphate (ADP) by glycerol kinase and ATP. Glycerol-3-phosphate (G3P) is then converted by glycerol phosphate dehydrogenase (GPO) to dihydroxyacetone phosphate (DAP) and hydrogen peroxide (H₂O₂).

In the last reaction, hydrogen peroxide (H₂O₂) reacts with 4-aminophenazone (4-AP) and p-chlorophenol in presence of peroxidase (POD) to give a red colored dye:



The intensity of the color formed is proportional to the triglycerides concentration in the sample^{1,2,3}.

CLINICAL SIGNIFICANCE

Triglycerides are fats that provide energy for the cell. Like cholesterol, they are delivered to the body's cells by lipoproteins in the blood. A diet with a lot of saturated fats or carbohydrates will raise the triglyceride levels. The increases in serum triglycerides are relatively non-specific. For example liver dysfunction resulting from hepatitis, extra hepatic biliary obstruction or cirrhosis, diabetes mellitus is associated with the increase^{3,6,7}.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

REAGENTS

R (Note 2)	GOOD pH 6.3	50 mmol/L
	p-Chlorophenol	2 mmol/L
	Lipoprotein lipase (LPL)	150000 U/L
	Glycerol kinase (GK)	500 U/L
	Glycerol-3-oxidasa (GPO)	3500 U/L
	Peroxidase (POD)	440 U/L
	4 - Aminophenazone (4-AP)	0,1 mmol/L
ATP	0,1 mmol/L	
TRIGLYCERIDES CAL	Aqueous primary standard	200 mg/dL

PREPARATION

Reagent and standard provided are ready to use.

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.

Do not use reagents over the expiration date.

Signs of reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 505 nm \geq 0,26.

ADDITIONAL EQUIPMENT

- Spectrophotometer or colorimeter measuring at 505 nm.
- Matched cuvettes 1,0 cm light path.
- General laboratory equipment.

SAMPLES

Serum or plasma¹.

Stability of the sample: 5 days at 2-8°C.

PROCEDURE

- Assay conditions:
 Wavelength: 505 nm (490-550)
 Cuvette: 1 cm light path
 Temperature: 37°C / 15-25°C
- Adjust the instrument to zero with distilled water.
- Pipette into a cuvette:

	Blank	Standard	Sample
R (mL)	1,0	1,0	1,0
Standard ^(Note 1,3,4) (μL)	--	10	--
Sample (μL)	--	--	10

- Mix and incubate for 5 min at 37°C or 10 min at 15-25°C.
- Read the absorbance (A) of the samples and standard, against the Blank. The colour is stable for at least 30 minutes.

CALCULATIONS

$$\frac{(A)\text{Sample} - (A)\text{Blank}}{(A)\text{Standard} - (A)\text{Blank}} \times \text{Standard conc.} = \text{mg/dL triglycerides in the sample}$$

Conversion factor: mg/dL x 0,0113= mmol/L.

QUALITY CONTROL

Control Sera are recommended to monitor the performance of assay procedures: CONTROL Normal and Pathologic.

If control values are found outside the defined range, check the instrument, reagent and calibration material.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES

Men	40 – 160 mg/dL
Women	35 – 135 mg/dL

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Measuring range: From detection limit 0,000 mg/dL to linearity limit 1600 mg/dL.

If the concentration is greater than linearity limit dilute 1/2 the sample with CIna 9 g/L and multiply the result by 2.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
	Mean (mg/dL)	SD	CV (%)	
Mean (mg/dL)	109	224	111	224
SD	0,64	1,01	3,74	7,90
CV (%)	0,58	0,45	3,38	3,52

Sensitivity: 1 mg/dL = 0,0013 (A).

Accuracy: Results obtained using Audit Diagnostics reagents (y) did not show systematic differences when compared with other commercial reagent (x). The results obtained using 50 samples were the following: Correlation coefficient (r): 0,99810.

Regression equation: y= 0,9178x - 0,5426

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

No interferences were observed with bilirubin < 170 μmol/L, hemoglobin < 10 g/L².

A list of drugs and other interfering substances with cholesterol determination has been reported^{4,5}.

NOTES

- TRIGLYCERIDES CAL: Proceed carefully with this product because due its nature it can get contaminated easily.
- LCF (Lipid Clearing Factor) is integrated in the reagent.
- Calibration with the aqueous Standard may cause a systematic error in automatic procedures. In these cases, it is recommended to use a serum Calibrator.
- Use clean disposable pipette tips for its dispensation.
- Audit Diagnostics has instruction sheets for several automatic analyzers.**

BIBLIOGRAPHY

- Buccolo G et al. Quantitative determination of serum triglycerides by use of enzymes. Clin Chem 1973; 19 (5): 476-482.
- Fossati P et al. Clin. Chem 1982; 28(10): 2077-2080.
- Kaplan A et al. Tryglycerides. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 437 and Lipids 1194-1206.
- Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
- Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
- Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
- Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.