



شرکت دلتا درمان پارت



دفتر مرکزی: تهران، میدان آرژانتین، خیابان الوند، خیابان سی و پنجم، پلاک ۱۳، طبقه پنجم

تلفن: ۸۸۸۵۶۴۱۰ ۸۸۸۵۶۳۸۵ ۸۸۷۷۰۶۵۸ ۸۸۷۷۳۶۶۰ ۸۸۷۷۵۶۵۶

فکس : ۸۸۸۵۶۴۰۳

ایمیل : info@delta-dp.ir
وب سایت www.delta-dp.ir

Chem	ACE	Sample Type	Serum		
Chemistry		Print Name	ACE		
Reaction Type	FIXED TIME	Rection Direction	Negative		
Pri. Wave	340	Sec. Wave	700		
Unit	U/L	incubation Time	0	Decimal	0
Blank Time		Reaction Time	30	60	
	Sample. VOL		Reagent. VOL		
Standard	15 uL	R1	150 uL		
Decreased		R2			
Increased		R3			

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	ADA	Sample Type	Serum		
Chemistry		Print Name	ADA		
Reaction Type	KINETIC	Rection Direction	Positive		
Pri. Wave	546	Sec. Wave	700		
Unit	U/L	incubation Time	0	Decimal	0
Blank Time		Reaction Time	12	30	
	Sample. VOL		Reagent. VOL		
Standard	3 uL	R1	100 uL		
Decreased		R2	50 uL		
Increased		R3			

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	ALBUMIN	Sample Type	Serum		
Chemistry		Print Name	ALBUMIN		
Reaction Type	END POINT	Rection Direction	Positive		
Pri. Wave	605	Sec. Wave	700		
Unit	g/dl	incubation Time	0	Decimal	0.1
Blank Time	-2	-1	Reaction Time	17	18
	Sample. VOL		Reagent. VOL		
Standard	2 uL	R1	200 uL		
Decreased		R2			
Increased		R3			

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	ALP	Sample Type	Serum		
Chemistry		Print Name	ALP		
Reaction Type	KINETIC	Rection Direction	Positive		
Pri. Wave	412	Sec. Wave	660		
Unit	U/L	incubation Time	18	Decimal	0
Blank Time		Reaction Time	6	24	
	Sample. VOL		Reagent. VOL		
Standard	3 uL	R1	160 uL		
Decreased		R2	40 uL		
Increased		R3			

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	ALT	Sample Type	Serum		
Chemistry		Print Name	GPT-(ALT)		
Reaction Type	KINETIC	Rection Direction	Negative		
Pri. Wave	340	Sec. Wave	412		
Unit	U/L	incubation Time	18	Decimal	0
Blank Time		Reaction Time	6	24	
	Sample. VOL		Reagent. VOL		
Standard	20 uL	R1	160 uL		
Decreased		R2	40 uL		
Increased		R3			

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	AMYLASE	Sample Type	Serum		
Chemistry		Print Name	AMYLASE		
Reaction Type	KINETIC	Rection Direction	Positive		
Pri. Wave	412	Sec. Wave	660		
Unit	U/L	incubation Time	0	Decimal	0
Blank Time		Reaction Time	6	24	
	Sample. VOL		Reagent. VOL		
Standard	4 uL	R1	200 uL		
Decreased		R2			
Increased		R3			

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	ASO		Sample Type	Serum	
Chemistry			Print Name	ASO	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	605		Sec. Wave		
Unit	IU/ml		incubation Time	18	Decimal 0
Blank Time			Reaction Time	17	18

	Sample. VOL		Reagent. VOL
Standard	3 uL	R1	200 uL
Decreased	uL	R2	50 uL
Increased	uL	R3	uL

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	AST		Sample Type	Serum	
Chemistry			Print Name	GOT-(AST)	
Reaction Type	KINETIC		Rection Direction	Negative	
Pri. Wave	340		Sec. Wave	412	
Unit	U/L		incubation Time	18	Decimal 0
Blank Time			Reaction Time	6	24

	Sample. VOL		Reagent. VOL
Standard	20 uL	R1	160 uL
Decreased	uL	R2	40 uL
Increased	uL	R3	uL

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	BILL-T		Sample Type	Serum	
Chemistry			Print Name	Total Bilirubin	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	546		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0.01
Blank Time	-2 -1		Reaction Time	29	30

	Sample. VOL		Reagent. VOL
Standard	5 uL	R1	160 uL
Decreased	uL	R2	40 uL
Increased	uL	R3	uL

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	BILL-D		Sample Type	Serum	
Chemistry			Print Name	Direct Bilirubin	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	546		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0.01
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	20	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

Chem	C3		Sample Type	Serum	
Chemistry			Print Name	C3	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	340		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	LogitLog 4P (6POINT)	
REPLICATES	2	

Chem	C4		Sample Type	Serum	
Chemistry			Print Name	C4	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	340		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0.1
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	4	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	LogitLog 4P (6POINT)	
REPLICATES	2	

Chem	CALCIUM	Sample Type	Serum		
Chemistry		Print Name	CALCIUM		
Reaction Type	END POINT	Rection Direction	Positive		
Pri. Wave	660	Sec. Wave	700		
Unit	mg/dl	incubation Time	0	Decimal	0.1
Blank Time	-2	-1	Reaction Time	17	18

	Sample. VOL		Reagent. VOL
Standard	2 uL	R1	200 uL
Decreased	uL	R2	uL
Increased	uL	R3	uL

CALIBRATION	
RULE	TWO POINT LINEAR
REPLICATES	2

Chem	CHOLESTEROL	Sample Type	Serum		
Chemistry		Print Name	CHOLESTEROL		
Reaction Type	END POINT	Rection Direction	Positive		
Pri. Wave	505	Sec. Wave	660		
Unit	mg/dl	incubation Time	0	Decimal	0
Blank Time		Reaction Time	58	59	

	Sample. VOL		Reagent. VOL
Standard	2 uL	R1	200 uL
Decreased	uL	R2	uL
Increased	uL	R3	uL

CALIBRATION	
RULE	TWO POINT LINEAR
REPLICATES	2

Chem	CK-NAC	Sample Type	Serum		
Chemistry		Print Name	CK-NAC		
Reaction Type	KINETIC	Rection Direction	Positive		
Pri. Wave	340	Sec. Wave	412		
Unit	U/L	incubation Time	18	Decimal	0
Blank Time		Reaction Time	12	30	

	Sample. VOL		Reagent. VOL
Standard	8 uL	R1	160 uL
Decreased	uL	R2	40 uL
Increased	uL	R3	uL

CALIBRATION	
RULE	TWO POINT LINEAR
REPLICATES	2

Chem	CK-MB		Sample Type	Serum	
Chemistry			Print Name	CK-MB	
Reaction Type	KINETIC		Rection Direction	Positive	
Pri. Wave	340		Sec. Wave	412	
Unit	U/L		incubation Time	18	Decimal 0.1
Blank Time			Reaction Time	12	30
	Sample. VOL		Reagent. VOL		
Standard	16	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	COPPER		Sample Type	Serum	
Chemistry			Print Name	COPPER	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	570		Sec. Wave	700	
Unit	µg/dl		incubation Time	0	Decimal 0
Blank Time	-2	-1	Reaction Time	17	18
	Sample. VOL		Reagent. VOL		
Standard	10	uL	R1	200	uL
Decreased		uL	R2		uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	CRP		Sample Type	Serum	
Chemistry			Print Name	CRP	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	605		Sec. Wave		
Unit	mg/L		incubation Time	18	Decimal 0.1
Blank Time			Reaction Time	17	18
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	SPLINE (5POINT)				
REPLICATES	2				

CHem	Creatinine Jaffe		Sample Type	Serum	
CHemistry			Print Name	Creatinine Jaffe	
Reaction Type	FIXED TIME		Rection Direction	Positive	
Pri. Wave	505		Sec. Wave	660	
Unit	mg/dl		incubation Time	0	Decimal 0.01
Blank Time			Reaction Time	6	15
	Sample. VOL		Reagent. VOL		
Standard	20	uL	R1	100	uL
Decreased		uL	R2	100	uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

CHem	FERRITIN		Sample Type	Serum	
CHemistry			Print Name	FERRITIN	
Reaction Type	FIXED TIME		Rection Direction	Positive	
Pri. Wave	570		Sec. Wave	800	
Unit	µg/L		incubation Time	18	Decimal 0
Blank Time			Reaction Time	3	30
	Sample. VOL		Reagent. VOL		
Standard	10	uL	R1	120	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	LogitLog 4P (6POINT)	
REPLICATES	2	

CHem	GGT		Sample Type	Serum	
CHemistry			Print Name	GGT	
Reaction Type	KINETIC		Rection Direction	Positive	
Pri. Wave	412		Sec. Wave	660	
Unit	U/L		incubation Time	18	Decimal 0
Blank Time			Reaction Time	6	24
	Sample. VOL		Reagent. VOL		
Standard	20	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

Chem	GLUCOSE		Sample Type	Serum	
Chemistry			Print Name	GLUCOSE	
Reaction Type	END POINT		Reaction Direction	Positive	
Pri. Wave	505		Sec. Wave	660	
Unit	mg/dl		incubation Time	0	Decimal 0
Blank Time			Reaction Time	58	59
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	200	uL
Decreased		uL	R2		uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

Chem	HDL-C		Sample Type	Serum	
Chemistry			Print Name	HDL-C	
Reaction Type	END POINT		Reaction Direction	Positive	
Pri. Wave	605		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0.1
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	180	uL
Decreased		uL	R2	60	uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

Chem	IgA		Sample Type	Serum	
Chemistry			Print Name	IgA	
Reaction Type	END POINT		Reaction Direction	Positive	
Pri. Wave	605		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	LogitLog 4P (6POINT)	
REPLICATES	2	

CHem	IgG		Sample Type	Serum	
CHemistry			Print Name	IgG	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	605		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	240	uL
Decreased		uL	R2	60	uL
Increased		uL	R3		uL

CALIBRATION

RULE	LogitLog 4P (6POINT)	
REPLICATES	2	

CHem	IgM		Sample Type	Serum	
CHemistry			Print Name	IgM	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	340		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	LogitLog 4P (6POINT)	
REPLICATES	2	

CHem	IRON		Sample Type	Serum	
CHemistry			Print Name	IRON	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	570		Sec. Wave		
Unit	µg/dl		incubation Time	18	Decimal 0
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	20	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

Chem	LACTATE		Sample Type	Serum	
Chemistry			Print Name	LACTATE	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	505		Sec. Wave	660	
Unit	mg/dl		incubation Time	0	Decimal 0.1
Blank Time			Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	200	uL
Decreased		uL	R2		uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

Chem	LDH		Sample Type	Serum	
Chemistry			Print Name	LDH	
Reaction Type	KINETIC		Rection Direction	negative	
Pri. Wave	340		Sec. Wave	412	
Unit	U/L		incubation Time	18	Decimal 0
Blank Time			Reaction Time	6	24
	Sample. VOL		Reagent. VOL		
Standard	4	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

Chem	LDL-C		Sample Type	Serum	
Chemistry			Print Name	LDL-C	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	605		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	180	uL
Decreased		uL	R2	60	uL
Increased		uL	R3		uL

CALIBRATION

RULE	TWO POINT LINEAR	
REPLICATES	2	

Chem	LIPASE		Sample Type	Serum	
Chemistry			Print Name	LIPASE	
Reaction Type	KINETIC		Rection Direction	Positive	
Pri. Wave	570		Sec. Wave	700	
Unit	U/L		incubation Time	18	Decimal 0
Blank Time			Reaction Time	6	18
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	165	uL
Decreased		uL	R2	33	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	MAGNESIUM		Sample Type	Serum	
Chemistry			Print Name	MAGNESIUM	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	546		Sec. Wave	700	
Unit	mg/dl		incubation Time	0	Decimal 0.01
Blank Time	-2	-1	Reaction Time	17	18
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	200	uL
Decreased		uL	R2		uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	Microalbumin		Sample Type	Urine	
Chemistry			Print Name	Microalbumin	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	340		Sec. Wave	700	
Unit	mg/L		incubation Time	18	Decimal 0.1
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	4	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	LogitLog 4P (6POINT)				
REPLICATES	2				

CHem	PHOSPHORUS		Sample Type	Serum	
CHemistry			Print Name	PHOSPHORUS	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	340		Sec. Wave	700	
Unit	mg/dl		incubation Time	0	Decimal 0.1
Blank Time	-2	-1	Reaction Time	17	18

	Sample. VOL		Reagent. VOL
Standard	2 uL	R1	200 uL
Decreased	uL	R2	uL
Increased	uL	R3	uL

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

CHem	RF		Sample Type	Serum	
CHemistry			Print Name	RF	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	605		Sec. Wave		
Unit	IU/ml		incubation Time	18	Decimal 0.1
Blank Time			Reaction Time	29	30

	Sample. VOL		Reagent. VOL
Standard	4 uL	R1	160 uL
Decreased	uL	R2	40 uL
Increased	uL	R3	uL

CALIBRATION

RULE	SPLINE (6POINT)
REPLICATES	2

CHem	TOTAL PROTEIN		Sample Type	Serum	
CHemistry			Print Name	TOTAL PROTEIN	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	546		Sec. Wave	700	
Unit	g/dl		incubation Time	0	Decimal 0.1
Blank Time	-2	-1	Reaction Time	17	18

	Sample. VOL		Reagent. VOL
Standard	5 uL	R1	200 uL
Decreased	uL	R2	uL
Increased	uL	R3	uL

CALIBRATION

RULE	TWO POINT LINEAR
REPLICATES	2

Chem	TIBC		Sample Type	Serum	
Chemistry			Print Name	TIBC	
Reaction Type	END POINT		Reaction Direction	Negative	
Pri. Wave	660		Sec. Wave	800	
Unit	µg/dl		incubation Time	18	Decimal 0
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	20	uL	R1	200	uL
Decreased		uL	R2	60	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	TRIGLYCERIDES		Sample Type	Serum	
Chemistry			Print Name	TRIGLYCERIDES	
Reaction Type	END POINT		Reaction Direction	Positive	
Pri. Wave	505		Sec. Wave	660	
Unit	mg/dl		incubation Time	0	Decimal 0
Blank Time			Reaction Time	58	59
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	200	uL
Decreased		uL	R2		uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	UREA		Sample Type	Serum	
Chemistry			Print Name	UREA	
Reaction Type	FIXED TIME		Reaction Direction	Negative	
Pri. Wave	340		Sec. Wave	412	
Unit	mg/dl		incubation Time	18	Decimal 0.1
Blank Time			Reaction Time	6	15
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	URIC ACID		Sample Type	Serum	
Chemistry			Print Name	URIC ACID	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	546		Sec. Wave	700	
Unit	mg/dl		incubation Time	18	Decimal 0.1
Blank Time	-2 -1		Reaction Time	28	29
	Sample. VOL		Reagent. VOL		
Standard	8	uL	R1	160	uL
Decreased		uL	R2	160	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	Urine Protein		Sample Type	Serum	
Chemistry			Print Name	Urine Protein	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	505		Sec. Wave	700	
Unit	mg/L		incubation Time	18	Decimal 0.1
Blank Time	-2 -1		Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	16	uL	R1	200	uL
Decreased		uL	R2	100	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	SPLINE (6POINT)				
REPLICATES	2				

Chem	ZN		Sample Type	Serum	
Chemistry			Print Name	Zinc	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	570		Sec. Wave	700	
	µg/dl		incubation Time	0	Decimal 0
Blank Time	-2 -1		Reaction Time	17	18
	Sample. VOL		Reagent. VOL		
Standard	10	uL	R1	200	uL
Decreased		uL	R2		uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	CRP HL		Sample Type	Serum	
Chemistry			Print Name	CRP HL	
Reaction Type	FIXED TIME		Rection Direction	Positive	
Pri. Wave	605		Sec. Wave		
Unit	mg/dl		incubation Time	18	Decimal 0.1
Blank Time			Reaction Time	3	30
	Sample. VOL		Reagent. VOL		
Standard	2	uL	R1	160	uL
Decreased		uL	R2	40	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	SPLINE (6 POINTS)				
REPLICATES	2				

Chem	HB		Sample Type	Whole Blood	
Chemistry			Print Name	HB	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	505		Sec. Wave	800	
Unit	umol/L		incubation Time	18	Decimal 0.01
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	12	uL	R1	180	uL
Decreased		uL	R2	60	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				

Chem	A1c		Sample Type	Whole Blood	
Chemistry			Print Name	A1c	
Reaction Type	END POINT		Rection Direction	Positive	
Pri. Wave	660		Sec. Wave	800	
Unit	umol/L		incubation Time	18	Decimal 0.01
Blank Time	-2	-1	Reaction Time	29	30
	Sample. VOL		Reagent. VOL		
Standard	12	uL	R1	180	uL
Decreased		uL	R2	60	uL
Increased		uL	R3		uL
CALIBRATION					
RULE	TWO POINT LINEAR				
REPLICATES	2				